



Product Service

CERTIFICATE

No. Z1A 08 04 17955 017

Holder of Certificate: Omicron electronics GmbH
 Oberes Ried 1
 6833 Klaus
 AUSTRIA

Factory(ies): 17955

Certification Mark:



Product: Electronic measuring equipment
 (Measuring, testing and signal generating equipment)

Model(s): CMC 256plus, CMC 356

Parameters:

Rated voltage:	100-240V	
Rated frequency:	50/60Hz	
Rated input current:	12-10A	
Rated output:	CMC 256plus:	4x300V/6x12,5A
	CMC 356:	4x300V/6x32A
Protection Class:	I	

Tested according to: EN 61010-1:2001

The product meets the requirements of the German Equipment and Product Safety Act. The Certification marks shown above can be affixed on the product. The certification marks must not be altered in any way. The use of the GS-Mark is permitted until the listed date, the use of the TÜV-Mark is unlimited, unless it is cancelled. See also notes overleaf.

Test report no.: 028-71323900-100

GS-Mark valid until: 2013-04-07

Date, 2008-04-09

Page 1 of 1

