



Product Service

CERTIFICATE

No. Z1A 10 05 17955 022

Holder of Certificate: Omicron electronics GmbHOberes Ried 1
6833 Klaus
AUSTRIA**Factory(ies):** 17955**Certification Mark:****Product:** Electronic measuring equipment
(Measuring, testing and signal generating equipment)**Model(s):** CMC 256plus, CMC 356, CMC 353**Parameters:**
Rated voltage: 100-240V
Rated frequency: 50/60Hz
Rated input current: 12-10A
Rated output: CMC 256plus: 4x300V/6x12,5A
CMC 356: 4x300V/6x32A
CMC 353: 4x300V/3x32A
Protection Class: I**Tested according to:** EN 61010-1:2001

The product meets the safety and health requirements of the German Equipment and Product Safety Act section 7 subsection 1 sentence 2 GPSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. See also notes overleaf.

Test report no.: 028-71323900-200**Valid until:** 2015-05-17**Date,** 2010-05-20 (Theo Butz)

Page 1 of 1

