◆ ZERTIFIKAT ◆ CERTIFICATE ◆ CERTIFICAT ◆ CERTIFICADO ◆ CERTIFICATO ◆ CERTIFICAAT ◆ СВИДЕТЕЛЬСТВО ◆ 認証証書 ◆

Certificate

No: AL 96 05 17955 008

TÜV PRODUCT SERVICE

Omicron Electronics GmbH

Brolsstr. 39
A - 6844 Altach

with production facilities
17955

is authorized to label the following products with the
certification mark A,B or L
as shown in the certification mark list. See also notes overleaf.

Product:	**Netzgeräte** **(Voltage/current amplifier)**
Model:	**Typ / type CMS 156**

Parameters:

Nennspannung: (Rated input voltage)	110-240 V
Nennleistung: (Rated input power)	1000 VA
Nennfrequenz: (Rated frequency)	50/60 Hz
Schutzklasse: Protection class)	I

The product meets the requirements of the Equipment Safety Law and was tested according to:

EN 61010-1: 1993

Report No: 01210505501

Released with the above certificate number by the certification body of TÜV PRODUCT SERVICE GMBH.

Organization unit: 1ITE / abs
Date: 05-14-1996

Translation of the German original.